



Translation

4th Supplement

(Supplement in accordance with Directive 94/9/EC Annex III number 6)

to the EC-Type Examination Certificate BVS 03 ATEX E 406 X

Equipment: Gas detection transmitter type P3U

Manufacturer: Dräger Safety AG & Co. KGaA

Address: D-23560 Lübeck

Description

The Essential Health and Safety Requirements with respect to the measuring function for explosion protection are assured by application of:

EN 50104:2002 + A1:2004

EN 50271:2001

This supplement to the EC-type examination certificate covers operation of the devices with pump module or relay module and modifications of the software (main).

This supplement to the EC-type examination certificate covers devices type P3U with software versions 7.5, 7.6, 7.8 and 7.9 (main), V13 (SIOS) and V11, V12 and V13 (pump module) for data transmission via the 4-20 mA interface.

This supplement to the EC-type examination certificate covers the measuring function for oxygen (measurement of inertisation) in the measuring range 0 - 25 %(v/v).

Test report

Test report PFG-no. 413000504P NII dated 15/10/2010

Special conditions for safe use

- See 3rd supplement to the EC-type examination certificate BVS 03 ATEX E 406 X

DEKRA EXAM GmbH

Bochum, dated 15/10/2010

Signed: Simanski

Certification body

Signed: Dr. Kiesewetter

Special services unit

We confirm the correctness of the translation from the German original.
In the case of arbitration only the German wording shall be valid and binding.

44809 Bochum, 15. October 2010
PFG-Kie

DEKRA EXAM GmbH

A blue ink signature, appearing to read 'Kie', written over a horizontal line.

Certification body

A blue ink signature, appearing to read 'H. W. ...', written over a horizontal line.

Special services unit